

This is an Open Access document downloaded from ORCA, Cardiff University's institutional repository: <https://orca.cardiff.ac.uk/id/eprint/91689/>

This is the author's version of a work that was submitted to / accepted for publication.

Citation for final published version:

Huckle, D., Hall, J., Bowen, Jenna , Allender, Christopher, Cuenca, Jerome and Porch, Adrian 2014. Towards an efficient biosensor for the detection of lipopolysaccharide in sepsis using molecularly imprinted polymers. *British Journal of Anaesthesia* 112 (1) , p. 199. 10.1093/bja/aet373

Publishers page: <https://doi.org/10.1093/bja/aet373>

Please note:

Changes made as a result of publishing processes such as copy-editing, formatting and page numbers may not be reflected in this version. For the definitive version of this publication, please refer to the published source. You are advised to consult the publisher's version if you wish to cite this paper.

This version is being made available in accordance with publisher policies. See <http://orca.cf.ac.uk/policies.html> for usage policies. Copyright and moral rights for publications made available in ORCA are retained by the copyright holders.



Towards an efficient biosensor for the detection of lipopolysaccharide in sepsis using molecularly imprinted polymers

D. Huckle¹, J. Hall¹, J. Bowen², C. Allender², J. Cuenca³ and A. Porch³

¹Cardiff Institute of Infection and Immunity, School of Medicine,

²School of Pharmacy and Pharmaceutical Science and

³Cardiff School of Engineering, Cardiff University, Cardiff, UK

The true potential of circulating biomarker detection and surveillance in patients with sepsis is yet to be elucidated. The focus of this project is lipopolysaccharide (LPS), a major constituent of gram-negative bacteria outer cell walls, and its use as a sepsis marker. The overarching hypothesis is the detection of circulating LPS will facilitate the early diagnosis of sepsis.

Molecular imprinting describes the generation of synthetic, polymeric receptors through the polymerization of monomers around a template molecule (Fig. 12). The resultant polymers possess recognition properties akin to an antibody with the ability to recognize the original target molecule and related species. Efficient molecularly imprinted polymers rely on strong interactions between the target, in this case, LPS, and the functional monomers that form the polymer matrix. Polymyxin B (PMB), a peptide antibiotic, has high affinity for LPS and is utilized in this study as a functional monomer. LPS is not a suitable template for conventional imprinting techniques; therefore, techniques that circumvent problems associated with the imprinting of biological macromolecules are needed and modified approaches that localize LPS at a surface via chemical immobilization at a solid–solvent interface (surface imprinting on a solid support) or via self-assembly at a solvent–solvent interface (suspension polymerization/ microfluidic techniques) have been investigated.

Microfluidics (MF) describes the science and technology that manipulates very small (10⁻⁹ to 10⁻¹⁸ litre) volumes of fluid. A system that generates segmented flow within polytetrafluoroethylene (PTFE) tubing has been constructed, providing small pockets of an aqueous monomer solution travelling in an organic solvent continuous phase. Using a bespoke cavity resonator, the in situ polymerization of the ‘beads’ of aqueous monomer solution in this MF system has been achieved using microwaves. Furthermore, the process of polymerization can be monitored by sensing the changes in resonant frequency imposed by the changing monomer solution as it polymerizes. Continued troubleshooting of suspension polymerization and microfluidic processes forms the foundation of ongoing work as the successful imprinting of LPS in polymer beads with integrated PMB becomes a possibility.




Fig 12 Schematic representation of traditional molecular imprinting method. A solution phase complex of the template molecule (LPS) and functional monomers (polymerizable PMB and acrylamide) forms, which is 'locked' in place during polymerization. Following removal of the template species, an imprint of the template within the polymer matrix remains; this imprinted material demonstrates a natural affinity for the template molecule.